



**Testing &
Certification
Australia**

No.: **1029.0401**

Certificate of Conformance

Issued To

*AtCor Medical Pty Ltd
Suite 11
1059 - 1063 Victoria Road
West Ryde NSW 2114
Australia*

for

Product: SphygmoCor Pulse Wave Analysis
Model Number: MM3
Manufacturer: PWV Medical Pty Ltd
Country of Origin: Australia
ECRI Number: 18-107
ECRI Category: Simulators, Noninvasive Blood Pressure

The electrical equipment described above and in the reference report/s, complies with the requirements of

AS/NZS 3200.1.0:1998

Exclusions: Clause 36 (EMC to AS/NZS 3200.1.2) & 48 (Biocompatibility to ISO 10993-1)

(TCA Report No. 42191.3 and 41320 are reference documents for this certificate)

***Issued under the Rules of Certification
of Testing & Certification Australia***

Date of Issue 28/02/2005

This certificate indicates compliance with the requirements of type examination. Subject to exclusions and any additional regulation requirements, the product complies with the relevant electrical safety and/or electromagnetic compatibility Essential Principles and is suitable for inclusion in equipment management programs as described in AS/NZS 3551. The certificate is valid for a period of three years from the date of issue.